



# MASTER

## Notes

- 1 See drawing 65-2610RK-R10 for RS-485 wiring.
- 2 Each set of relay contacts rated for 5 amps 250 VAC resistive.

|   |                    |  |  |
|---|--------------------|--|--|
| DO NOT SCALE DRAWING                          |                    | All Dimensions in Inches Unless Otherwise Noted. |  |
| Debur   | Break Edges .010 R | MATERIAL: N/A                                    |  |
| Tolerances & Finishes Unless Otherwise Noted: |                    | FINISH: N/A                                      |  |
| .XX ± .010                                    | Angles ± 0° 30'    |  |  |
| .XXX ± .005                                   | Conc. .010 TIR     |  |  |
| Fractions ± .015                              | Finish 125         |  |  |
| Finish for O-ring Grooves, 3 Sides            |                    | SCALE: None                                      |  |
|   |                    | APPROVED BY:                                     |  |
|   |                    | DATE: 02/01/10                                   |  |
|   |                    | PAGE 1 OF 1                                      |  |



**RKI Instruments, Inc.**  
www.rkiinstruments.com

**TITLE: Wiring Diagram , M2, IR CO2 to Beacon 200**

**A** DRAWN BY: GD PART NUMBER: 65-2630RK-XX-R03 REVISION: P1

IR CO2 Detector

Black

Green

Red

White