



Notes

- 1 See drawing 65-2610RK-R10 for RS-485 wiring.
- 2 Each set of relay contacts rated for 5 amps 250 VAC resistive.

MASTER

CO Detector,
Factory Wired

DO NOT SCALE DRAWING		All Dimensions In Inches Unless Otherwise Noted.	
Debur	Break Edges .010 R	MATERIAL: N/A	
Tolerances & Finishes Unless Otherwise Noted:		FINISH: N/A	
.XX ± .010		Angles ± 0° 30'	
.XXX ± .005		Conc. .010 TIR	
Fractions ± .015		Finish 125	
Finish for O-ring Grooves, 3 Sides		32	
		SCALE: None	

		RKI Instruments, Inc. www.rkiinstruments.com	
TITLE: Wiring Diagram, M2, NON XP CO to Beacon 200			
A	DRAWN BY: GD	PART NUMBER: 65-2633RK-R03	REVISION: P1
APPROVED BY: 		DATE: 02/05/10	
		PAGE 1 OF 1	

RKI Instruments, Inc.
 www.rkiinstruments.com

TITLE: Wiring Diagram, M2, NON XP CO to Beacon 200